




The people with the solutions that protect you and your environment everyday

Gastec Tube Datasheet

Hydrogen Chloride

HCl

No. **GAS14L**



Performance

Measuring Range	0.2 to 1 ppm	1 to 20 ppm	20 to 76 ppm
Number of Pump Strokes	2 to 5	1	1/2
Correction Factor	1/2 to 1/5	1	3.8
Sampling Time	1 minute per pump stroke		
Detecting Limit	0.05 ppm (n=5)		
Colour Change	Yellow → Pink		
Reaction Principle	Hydrogen Chloride reacts with indicator to produce pink stain. HCl + Indicator → Chemical reaction compound		
Coefficient of Variation	10% (for 1 to 5 ppm), 5% (for 5 to 20 ppm)		
Shelf Life	Up to 3 Years		
Corrections for temperature & humidity	Unnecessary		
Store the tubes at cool and dark place.			

Possible coexisting substances and their interferences

Substance	Concentration	Interference	Change colour by itself
Chlorine	$\geq 1/2$	Plus error	Produces pink stain
Nitric Acid	$\geq 1/10$	Plus error	Produces pink stain
Sulphur Dioxide	≥ 10 times	Plus error	Produces pink stain
Nitrogen Dioxide	≥ 5 times	Plus error	Produces pink stain
Hydrogen Fluoride	≥ 600 ppm	Plus error	Produces pink stain

Water Vapour and up to 600ppm of Hydrogen Fluoride are trapped in the pre-treatment (white) layer

Calibration gas generation

High pressure gas cylinder method

a1-cbiss Ltd

5 Valiant Way, Lairdside Technology Park, Tranmere, CH41 9HS

www.a1-cbiss.com | sales@a1-cbiss.com | Tel: +44(0)151 666 8300 | Fax: +44(0)151 666 8329

VAT No: 539 5178 13 Company Reg No: 2414633



Certificate No 996
ISO 9001, ISO 14001
OHSAS 18001