




The people with the solutions that protect you and your environment everyday

Gastec Tube Datasheet

Arsine

AsH₃

NO.GAS19LA



Performance

Measuring Range	0.04 to 0.1ppm	0.1 to 1.5ppm	1.5 to 2.4 ppm	2.4 to 10 ppm
Number of Pump Stroke	10	5	3	1
Correction Factor	0.4	1	1.6	6.7
Sampling Time	1.5 minutes per pump stroke			
Detecting Limit	0.02 ppm (n=10)			
Colour Change	Yellow → Red			
Reaction Formula	<p>Arsine reacts with Mercuric Chloride to produce Hydrogen Chloride. The indicator of the tube turns to red. $AsH_3 + 3HgCl_2 \rightarrow As(HgCl)_3 + 3HCl$ $HCl + \text{Basic compound} \rightarrow \text{Chloride compounds}$</p>			
Coefficient of Variation	10% (for 0.1 to 0.5 ppm), 5% (for 0.5 to 1.5 ppm)			
Shelf Life	Up to 2 Years			
Corrections for temperature & humidity	Unnecessary			
Store the tubes at cool and dark place.				

Possible coexisting substances and their interferences

Substance	Concentration	Interference	Change colour by itself
Hydrogen Chloride	-	No effect	No stain by itself
Diborane	-	No effect	No stain by itself
Phosphine	≥1/10 time	Plus error	Discolours red stain

Calibration gas generation High pressure gas cylinder method

a1-cbiss Ltd

5 Valiant Way, Lairdside Technology Park, Tranmere, CH41 9HS

www.a1-cbiss.com | sales@a1-cbiss.com | Tel: +44(0)151 666 8300 | Fax: +44(0)151 666 8329

VAT No: 539 5178 13 Company Reg No: 2414633



Certificate No 996
ISO 9001, ISO 14001
OHSAS 18001